



Certificate of Compliance

Certificate: 1985108 (LR 44340)

Master Contract: 162029

Project: 1985108

Date Issued: 2007/12/04

Issued to: Fluke Corporation

P.O. Box 9090
Everett, WA 98206-9090
USA
Attention: Mr. Thomas Smith

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Cindy Johnson, ASCT.

Authorized by: Ken Rutledge, Product Group Manager

PRODUCTS

CLASS 3631 05 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT
CLASS 3631 85 - ELECTRICAL EQUIPMENT FOR MEASUREMENT USE - Certified to US Standards

Defibrillator/Pacer Analyzers, Models Impulse 7000DP and Impulse 6000D, rated 15Vdc, 1.5A.

Note: The above model is Equipment Class 2, Pollution Degree 2, Installation (Measurement) Category I.

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1985108 (LR 44340)

Master Contract: 162029

Project: 1985108

Date Issued: 2007/12/04

CONDITIONS FOR CONSIDERATION

The equipment is supplied with an approved power supply cord set or power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

ISA-82.02.01 - Safety Standards for Electrical and Electronic Test, Measuring, Controlling and Related Equipment – General Requirements

Reference Standards

IEC Standard No. 61010-1:2001 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements